

BR

<b>Notice of Allowability</b>	<b>Application No.</b>	<b>Applicant(s)</b>	
	10/648,827	MIHASHI ET AL	
	Examiner	Art Unit	
	Scott J. Sugarman	2873	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTO-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1.  This communication is responsive to the amendment of 20 November 2006.
2.  The allowed claim(s) is/are 1-28.
3.  Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
  - a)  All b)  Some\* c)  None of the:
    1.  Certified copies of the priority documents have been received.
    2.  Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
    3.  Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\* Certified copies not received: \_\_\_\_\_.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.  
**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**

4.  A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
5.  CORRECTED DRAWINGS ( as "replacement sheets") must be submitted.
  - (a)  including changes required by the Notice of Draftsperson's Patent Drawing Review ( PTO-948) attached
    - 1)  hereto or 2)  to Paper No./Mail Date \_\_\_\_\_.
  - (b)  including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_.

Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6.  DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

#### Attachment(s)

1.  Notice of References Cited (PTO-892)
2.  Notice of Draftsperson's Patent Drawing Review (PTO-948)
3.  Information Disclosure Statements (PTO/SB/08),  
Paper No./Mail Date \_\_\_\_\_
4.  Examiner's Comment Regarding Requirement for Deposit  
of Biological Material
5.  Notice of Informal Patent Application
6.  Interview Summary (PTO-413),  
Paper No./Mail Date \_\_\_\_\_.
7.  Examiner's Amendment/Comment
8.  Examiner's Statement of Reasons for Allowance
9.  Other \_\_\_\_\_.

## DETAILED ACTION

### ***Allowable Subject Matter***

Claims 1-28 are allowed.

The following is an examiner's statement of reasons for allowance:

The prior art does not teach or suggest an eye characteristic measuring apparatus having a light source part that emits a light flux with a first wavelength; a first illuminating optical system configured to illuminate a minute area on a retina of a subject eye with the light flux from the light source part; a first light receiving optical system configured to receive a part of a reflected light flux reflected and returned from the retina of the subject eye through a first conversion member including a lens part having a long focal point or first sensitivity and configured to convert the reflected light flux into substantially at least 17 beams; a second light receiving optical system configured to receive a part of the reflected light flux reflected and returned from the retina of the subject eye through a second conversion member including a lens part having a short focal point or second sensitivity and configured to convert the reflected light flux into substantially at least 17 beams, the first sensitivity being greater than the second sensitivity; a first light receiving part configured to receive the received light flux of the first light receiving optical system; a second light receiving part configured to receive the received light flux of the second light receiving optical system; a measurement condition setting part configured to set a measurement condition of the first and/or the second light receiving

optical system on the basis of an output signal from the second light receiving part; and an arithmetic part configured to obtain an optical characteristic of the subject eye on the basis of an output of the first light receiving part and/or the second light receiving part (independent claims 1 and 3).

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

### ***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Scott J. Sugarman whose telephone number is (571)272-2340.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Ricky L. Mack can be reached on (571)272-2333. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



Scott J. Sugarman  
Primary Examiner  
Art Unit 2873

sjs  
March 18, 2007